


Patient's name: _____

Social insurance number: _____

BariSling has been designed for lifting persons weighing 200 kg or more. It can be used in combination with mobile or stationary (overhead) lifts, with a four-point slingbar or with a BariTurn. BariSling has six suspension points. It has two slingbar loops at the back and four below, under the legs, to distribute pressure comfortably and safely.



Max: 570 kg



Functional inspection

Visual and mechanical inspection

Check the condition and function of the sling regularly. Always inspect the product after laundering. Check to ensure that seams and material are free from damage. Check to ensure that fabric is not worn or faded. Apply load to the device and check to ensure that clasps, handles, etc. withstand heavy load. If there are signs of wear, the product must be discarded.

Accessories

SoftLegSupport

StrapsPadding

Which lift/slingbar is used?

Stationary lift unit with four-point/sling crossbar

SWL: _____

Mobile lift with four-point/sling crossbar

SWL: _____

BariTurn (2 x lift units)

Slingbar width: _____

SWL: _____

Note! BariSling is recommended for use with four-point/sling crossbar or BariTurn!

Which loop alternatives are used?

Instructions for use

Can the sling be left under the patient in the wheelchair?

Yes No

Have staff that use the assistive device:

Read the instructions?

Yes No

Received practical instruction in patient transfers?

Yes No

In which situations will the BariSling be used?

Instructions to personnel:

e.g.; Raise the bed side rail on the side towards which the patient is to be turned. Placement of head/ arms/hands?
Placement of legs/feet? Use bed functions.

Number of caregivers: _____

Instructions to patient:

e.g.; In what way can the patient participate actively in the transfer situation? Which instructions are used?

Other:

Trial-fitted by: _____

Date: _____

Phone: _____

Phone: _____

Handicare AB
Veddestav. 15, Box 640
SE-175 27 Järfälla
SWEDEN

Tel: +46 (0)8-557 62 200
Fax: +46 (0)8-557 62 299
E-mail: info@handicare.se

Internet: www.handicare.com

**In the events of changes or questions,
contact the trial-fitter.**

Manuals and instructions for functional testing are
available on Handicare's website www.handicare.com

