


Patient's name: _____

Social insurance number: _____

BariVest is a lift vest that has been design as an aid to gait training for patients weighing 180 kg or more. It can be used with both mobile and stationary lifts, with a four-point slingbar or with BariTurn. BariVest features a back section, a front section and a narrower section in the middle to enable comfortable fit in the crotch area. It has two slingbar loops, both front and back, as well as belts at three levels.



Functional inspection

Visual and mechanical inspection

Check the condition and function of the sling regularly. Always inspect the product after laundering. Check to ensure that seams and material are free from damage. Check to ensure that fabric is not worn or faded. Apply load to the device and check to ensure that clasps, handles, etc. withstand heavy load. If there are signs of wear, the product must be discarded.

Which size is used?

XXX-Large

XXXX-Large

Accessories

BariVestCover

StrapsPadding

Which lift/slingbar is used?

Stationary lift unit with four-point/sling crossbar

SWL: _____

Mobile lift with four-point/sling crossbar

SWL: _____

BariTurn (2 x lift units)

Slingbar width: _____

SWL: _____

Note! BariVest is recommended for use with four-point/sling crossbar or BariTurn!

Which loop alternatives are used?

Instructions for use

Can the sling be left under the patient in the wheelchair?

Yes No

Have staff that use the assistive device:

Read the instructions?

Yes No

Received practical instruction in patient transfers?

Yes No

In which situations will the BariVest be used?

Instructions to personnel:

e.g.; Raise the bed side rail on the side towards which the patient is to be turned. Placement of head/ arms/hands?
Placement of legs/feet? Use bed functions.

Number of caregivers: _____

Instructions to patient:

e.g.; In what way can the patient participate actively in the transfer situation? Which instructions are used?

Other:

Trial-fitted by:

Date: _____

Phone: _____

Phone: _____

Handicare AB
Veddestav. 15, Box 640
SE-175 27 Järfälla
SWEDEN

Tel: +46 (0)8-557 62 200
Fax: +46 (0)8-557 62 299
E-mail: info@handicare.se

Internet: www.handicare.com

**In the events of changes or questions,
contact the trial-fitter.**

Manuals and instructions for functional testing are
available on Handicare's website www.handicare.com

