



Product Service

## CE Test Report

Spirit series  
Hospital Beds

Testing performed for

**CHG**<sup>®</sup>  
HOSPITAL BEDS

Report: KI1007332-100

Issued: 2012-12-18



Client: CHG Hospital Beds Inc. (82715)  
1020 Adelaide Street South, London, ON, N6E 1R6, Canada  
Contact: Paul Mitchener

Manufacturing place: CHG Hospital Beds Inc. (82715)  
1020 Adelaide Street South, London, ON, N6E 1R6, Canada

Test subject: Product:  
Hospital Bed  
  
Type:  
Spirit series, Spirit Plus series, Spirit Select series

Test specification: **EN 60601-1:2006**, excluding Usability (clauses 7.1.1 and 12.2),  
Biocompatibility (clause 11.7), PEMS (clause 14), and EMC  
(clauses 7.9.2.2 and 17)

**EN 60601-2-52/AC:2011**, excluding usability (clause 201.12.2)

Purpose of examination: Test according to the test specification.

Test result: The test specifications are met

This technical report may only be quoted in full. Any use for advertising purposes must be granted in writing. This report is the result of a single examination of the object in question and is not generally applicable evaluation of the quality of other products in regular production.

This report does not constitute a certification or endorsement by TÜV SÜD.